

Certificate of Analysis


No. CA 18075

Product: Eichrom's Nickel Resin
Lot #: NIA052318
Date of Manufacture: 23-MAY-2018

Eichrom's Nickel Resin Lot # NIA052318 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	96.9 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %



Jill Bryant
Quality System Coordinator

May 29, 2018