

Certificate of Analysis

No. CA 18099

Product: Eichrom's Cation Exchange Resin Cartridges
 Date of Manufacture: 25-JUN-2018
 Lot #: C8FR18G

Eichrom's Cation Exchange resin cartridge Lot C8FR18G has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 033016C8F, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	Tan to orange-brown	Pass
Extractables	< 1 mg(residue) / g(resin)	0.1 mg/g
pH	≥ 4.0	pass
Wet Vol. Capacity	Minimum 1.7 meq/mL	1.9

Jill Bryant

 Jill Bryant
 Quality System Coordinator

June 26, 2018