



# Certificate of Analysis

No. CA 18100

Product: Eichrom's Cation Exchange Resin Cartridges; 50Wx8, 100-200 mesh,  
Hydrogen Form

Lot: C8MR18A

Date of Manufacture: 26-JUN-2018

Eichrom's Cation Exchange resin cartridge Lot C8MR18A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 060315C8M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg (residue) /g (resin)	n.d.
pH	> 4.0	pass
Wet Volume Capacity	min. 1.7 meq/mL	2.0 meq/mL (Per vendor cert)

Jill Bryant  
Quality System Coordinator

June 27, 2018