



A GCI COMPANY

**Certificate of Analysis**  
No. CA18148

Product: Eichrom's TEVA Resin Cartridges  
Lot #: TESR18F  
Date of Manufacture: 29-AUG-2018

Eichrom's TEVA Resin cartridge Lot # TESR18F has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, TES071118, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	435.5
% <sup>230</sup> Th stripped in 6M HCl		
Least acceptable mean (4 rep)	93.0	99.7
Max. standard deviation	2.5	0.1
% <sup>230</sup> Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (4 rep)	2.0	0.2
Max. standard deviation	1.5	0.1
% <sup>239</sup> Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (4 rep)	90.0	99.3
Max. standard deviation	2.5	0.3
% <sup>239</sup> Pu stripped in 6M HCl		
Maximum acceptable mean (4 rep)	1.0	0.0
Max. standard deviation	0.5	0.0
Extractant distribution ratio, <sup>230</sup> Th in 3M HNO <sub>3</sub>	>250	267
Extractant distribution ratio, <sup>230</sup> Th in 6M HCl	<0.01	0.000

Jill Bryant  
Quality System Coordinator

August 29, 2018

Eichrom Technologies LLC  
1955 University Lane, Lisle, IL 60532, USA  
www.eichrom.com

npo

eichrom