

# Certificate of Analysis

No. CA21032

Product: Eichrom's MnO<sub>2</sub> Resin Cartridges  
Lot #: MNAR21A  
Date of Manufacture: 1-FEB-2021

Eichrom's MnO<sub>2</sub> Resin cartridge Lot MNAR21A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, # MNA042320, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (mean, 4 reps)	> 95%	101.5 %
Ba-133 in Load solution (mean, 4 reps)	<1.0%	0.0 %

Jill Bryant  
Quality System Coordinator



February 1, 2021