



Certificate of Analysis

No. CA 21241

Product: Eichrom's Anion Exchange Resin
Lot #: 091721A4M
Date of Manufacture: 17-SEP-2021

Eichrom's Anion Exchange resin, Lot # 091721A4M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|---------------------|------------------------------|---------------|
| Appearance | White to light Yellow | Pass |
| Extractables | < 1 mg (residue) / g (resin) | n.d |
| Wet Volume Capacity | ≥ 1.0 meq/mL | 1.2 |



Jill Bryant
Quality System Coordinator

October 13, 2021