

# Certificate of Analysis

No. CA 24153

Product: Eichrom's Pb Resin  
 Lot #: PBS070824  
 Date of Manufacture: 8-JUL-2024

Eichrom's Pb Resin Lot # PBS070824, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.0 %
% Lead in Strip	> 99 %	100.0 %
Gravimetric Yield	80-120 %	88.6 %
Chemical Yield	80-120 %	88.5 %

  
 Jose Munoz  
 Quality Analyst

July 18, 2024